

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
 Data File : VU061314.D
 Acq On : 31 Oct 2024 16:15
 Operator : MD/SY
 Sample : VIBLK153
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK153

Quant Time: Nov 01 00:00:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 31 23:55:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	167395	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	153744	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	69156	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	58000	5.101	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.000%
7) Chloroethane-d5	1.911	69	54685	5.246	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.000%
11) 1,1-Dichloroethene-d2	2.560	65	24828	5.103	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.000%
20) 2-Butanone-d5	4.628	46	122996	53.275	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.560%
24) Chloroform-d	5.055	84	112000	5.189	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.800%
26) 1,2-Dichloroethane-d4	5.695	65	58157	5.406	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.200%
32) Benzene-d6	5.721	84	232910	5.376	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%
36) 1,2-Dichloropropane-d6	6.682	67	71217	5.468	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.400%
41) Toluene-d8	7.891	98	214497	5.295	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.000%
43) trans-1,3-Dichloroprop...	8.174	79	25510	4.798	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.000%
46) 2-Hexanone-d5	8.627	63	108040	49.617	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.240%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	50271	5.104	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	66208	5.406	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
Data File : VU061314.D
Acq On : 31 Oct 2024 16:15
Operator : MD/SY
Sample : VIBLK153
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK153

Quant Time: Nov 01 00:00:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 31 23:55:38 2024
Response via : Initial Calibration

