

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
 Data File : VU061312.D
 Acq On : 31 Oct 2024 15:26
 Operator : MD/SY
 Sample : VIBLK152
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK152

Quant Time: Nov 01 00:00:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 31 23:55:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	169008	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	153859	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	71210	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	54944	4.786	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.800%
7) Chloroethane-d5	1.911	69	51484	4.892	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.800%
11) 1,1-Dichloroethene-d2	2.560	65	24460	4.980	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.600%
20) 2-Butanone-d5	4.628	46	117031	50.208	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.420%
24) Chloroform-d	5.055	84	106857	4.903	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.695	65	54830	5.048	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.000%
32) Benzene-d6	5.718	84	225355	5.198	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
36) 1,2-Dichloropropane-d6	6.682	67	67295	5.163	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.200%
41) Toluene-d8	7.891	98	203877	5.030	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.600%
43) trans-1,3-Dichloroprop...	8.174	79	25233	4.743	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%
46) 2-Hexanone-d5	8.627	63	104322	47.874	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.740%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	48480	4.919	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	63968	5.073	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
Data File : VU061312.D
Acq On : 31 Oct 2024 15:26
Operator : MD/SY
Sample : VIBLK152
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK152

Quant Time: Nov 01 00:00:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 31 23:55:38 2024
Response via : Initial Calibration

