

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
 Data File : VU061316.D
 Acq On : 31 Oct 2024 17:04
 Operator : MD/SY
 Sample : P4629-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP21

Quant Time: Nov 01 00:01:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 31 23:55:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	173659	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	159975	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	70449	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	46691	3.958	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.200%
7) Chloroethane-d5	1.907	69	48569	4.491	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.560	65	22194	4.397	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.000%
20) 2-Butanone-d5	4.624	46	120024	50.113	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.220%
24) Chloroform-d	5.055	84	104649	4.674	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.695	65	56289	5.043	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.721	84	217942	4.835	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.682	67	67375	4.972	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.400%
41) Toluene-d8	7.891	98	197253	4.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	8.174	79	24535	4.435	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.800%
46) 2-Hexanone-d5	8.627	63	105710	46.656	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.320%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	49677	4.847	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	63395	5.082	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
Data File : VU061316.D
Acq On : 31 Oct 2024 17:04
Operator : MD/SY
Sample : P4629-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP21

Quant Time: Nov 01 00:01:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 31 23:55:38 2024
Response via : Initial Calibration

