

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
 Data File : VU061327.D
 Acq On : 31 Oct 2024 22:36
 Operator : MD/SY
 Sample : P4629-14
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP27

Quant Time: Nov 01 01:50:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	163393	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	151366	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	65671	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	46277	4.169	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.400%
7) Chloroethane-d5	1.907	69	49145	4.830	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.560	65	21806	4.592	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.800%
20) 2-Butanone-d5	4.624	46	120849	53.627	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.260%
24) Chloroform-d	5.055	84	107243	5.090	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	5.695	65	55660	5.300	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.000%
32) Benzene-d6	5.718	84	224452	5.263	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.200%
36) 1,2-Dichloropropane-d6	6.682	67	69986	5.458	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.200%
41) Toluene-d8	7.891	98	202611	5.081	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
43) trans-1,3-Dichloroprop...	8.174	79	23924	4.571	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.627	63	105297	49.117	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.240%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	49858	5.142	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	66244	5.696	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103124\
Data File : VU061327.D
Acq On : 31 Oct 2024 22:36
Operator : MD/SY
Sample : P4629-14
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP27

Quant Time: Nov 01 01:50:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

