

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061353.D
 Acq On : 01 Nov 2024 11:21
 Operator : MD/SY
 Sample : P4629-20
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Nov 04 04:00:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	181216	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	165830	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	74634	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	43306	3.518	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.400%	
7) Chloroethane-d5	1.911	69	47886	4.243	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.800%	
11) 1,1-Dichloroethene-d2	2.560	65	21601	4.101	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.000%	
20) 2-Butanone-d5	4.624	46	114497	45.812	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.620%	
24) Chloroform-d	5.055	84	105021	4.495	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.800%	
26) 1,2-Dichloroethane-d4	5.695	65	55227	4.742	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.800%	
32) Benzene-d6	5.721	84	212560	4.549	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.000%	
36) 1,2-Dichloropropane-d6	6.682	67	65364	4.653	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.000%	
41) Toluene-d8	7.891	98	194047	4.441	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.800%	
43) trans-1,3-Dichloroprop...	8.174	79	22701	3.959	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.200%	
46) 2-Hexanone-d5	8.627	63	98145	41.788	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	83.580%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	48584	4.573	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	62288	4.713	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.200%	
Target Compounds						
13) Acetone	2.647	43	7110	4.491	ug/L	98
25) Chloroform	5.081	83	30253	1.288	ug/L	99
35) Methylcyclohexane	6.685	83	15108	0.680	ug/L #	21
37) 1,2-Dichloropropane	6.682	63	7024	0.540	ug/L #	89
61) 1,2,3-Trichloropropane	11.804	75	9910	1.333	ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061353.D
Acq On : 01 Nov 2024 11:21
Operator : MD/SY
Sample : P4629-20
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Nov 04 04:00:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

