

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061363.D
 Acq On : 01 Nov 2024 15:26
 Operator : MD/SY
 Sample : P4625-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Nov 04 04:01:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	143758	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	131782	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	58043	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	37750	3.866	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.400%	
7) Chloroethane-d5	1.907	69	42268	4.721	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.400%	
11) 1,1-Dichloroethene-d2	2.560	65	18164	4.347	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	87.000%	
20) 2-Butanone-d5	4.624	46	103074	51.987	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	103.980%	
24) Chloroform-d	5.052	84	91141	4.917	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
26) 1,2-Dichloroethane-d4	5.695	65	47822	5.176	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.600%	
32) Benzene-d6	5.718	84	185684	5.001	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.000%	
36) 1,2-Dichloropropane-d6	6.682	67	57472	5.148	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.000%	
41) Toluene-d8	7.891	98	166320	4.790	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.800%	
43) trans-1,3-Dichloroprop...	8.177	79	20286	4.452	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.000%	
46) 2-Hexanone-d5	8.627	63	88401	47.364	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.720%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	42943	5.087	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.800%	
66) 1,2-Dichlorobenzene-d4	12.187	152	53950	5.249	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	
Target Compounds						
					Qvalue	
13) Acetone	2.644	43	6325	5.037	ug/L	95
16) Methylene chloride	3.039	84	11996	1.198	ug/L	98
35) Methylcyclohexane	6.682	83	13573	0.769	ug/L #	21
37) 1,2-Dichloropropane	6.679	63	6341	0.614	ug/L #	89
42) Toluene	7.965	91	5322	0.120	ug/L	97
61) 1,2,3-Trichloropropane	11.804	75	7514	1.300	ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061363.D
Acq On : 01 Nov 2024 15:26
Operator : MD/SY
Sample : P4625-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Nov 04 04:01:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

