

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061368.D
 Acq On : 01 Nov 2024 17:29
 Operator : MD/SY
 Sample : P4625-02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Nov 04 04:01:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	161685	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	150201	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	65132	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	38129	3.472	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	69.400%	
7) Chloroethane-d5	1.907	69	36881	3.663	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	73.200%	
11) 1,1-Dichloroethene-d2	2.560	65	18097	3.851	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	77.000%	
20) 2-Butanone-d5	4.621	46	104293	46.770	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.540%	
24) Chloroform-d	5.055	84	92685	4.446	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
26) 1,2-Dichloroethane-d4	5.695	65	49957	4.808	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.200%	
32) Benzene-d6	5.718	84	186897	4.416	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.400%	
36) 1,2-Dichloropropane-d6	6.682	67	59094	4.644	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.800%	
41) Toluene-d8	7.891	98	169216	4.276	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.600%	
43) trans-1,3-Dichloroprop...	8.177	79	20012	3.853	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.000%	
46) 2-Hexanone-d5	8.627	63	90223	42.412	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.820%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	43153	4.485	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	54417	4.718	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.400%	
Target Compounds						Qvalue
35) Methylcyclohexane	6.682	83	13010	0.647	ug/L #	21
37) 1,2-Dichloropropane	6.682	63	6353	0.540	ug/L #	89
61) 1,2,3-Trichloropropane	11.804	75	8279	1.276	ug/L #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061368.D
Acq On : 01 Nov 2024 17:29
Operator : MD/SY
Sample : P4625-02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Nov 04 04:01:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

