

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061387.D
 Acq On : 02 Nov 2024 02:04
 Operator : MD/SY
 Sample : P4671-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Nov 04 05:09:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	159776	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	151353	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	64901	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	17180	1.583	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	31.600%#	
7) Chloroethane-d5	1.907	69	19437	1.953	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	39.000%#	
11) 1,1-Dichloroethene-d2	2.560	65	8049	1.733	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	34.600%#	
20) 2-Butanone-d5	4.631	46	49843	22.619	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	45.240%	
24) Chloroform-d	5.055	84	42770	2.076	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	41.600%#	
26) 1,2-Dichloroethane-d4	5.698	65	22557	2.197	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	44.000%#	
32) Benzene-d6	5.721	84	84166	1.974	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	39.400%#	
36) 1,2-Dichloropropane-d6	6.682	67	26873	2.096	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	42.000%#	
41) Toluene-d8	7.891	98	75726	1.899	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	38.000%#	
43) trans-1,3-Dichloroprop...	8.177	79	8366	1.598	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	32.000%#	
46) 2-Hexanone-d5	8.631	63	41362	19.295	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	38.600%#	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	21455	2.213	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	44.200%#	
66) 1,2-Dichlorobenzene-d4	12.187	152	26347	2.293	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	45.800%#	
Target Compounds						
					Qvalue	
35) Methylcyclohexane	6.679	83	6344	0.313	ug/L #	21
37) 1,2-Dichloropropane	6.682	63	3057	0.258	ug/L #	89
61) 1,2,3-Trichloropropane	11.804	75	7871	1.217	ug/L #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061387.D
Acq On : 02 Nov 2024 02:04
Operator : MD/SY
Sample : P4671-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Nov 04 05:09:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

