

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061393.D
 Acq On : 02 Nov 2024 04:31
 Operator : MD/SY
 Sample : P4668-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Nov 04 05:10:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	168431	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	156196	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	68160	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	34339	3.001	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.000%
7) Chloroethane-d5	1.907	69	38616	3.682	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.600%
11) 1,1-Dichloroethene-d2	2.560	65	17207	3.515	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.400%
20) 2-Butanone-d5	4.624	46	105953	45.611	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.220%
24) Chloroform-d	5.052	84	87845	4.045	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.800%
26) 1,2-Dichloroethane-d4	5.695	65	46173	4.265	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.400%
32) Benzene-d6	5.717	84	178269	4.051	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.000%
36) 1,2-Dichloropropane-d6	6.682	67	56703	4.285	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.800%
41) Toluene-d8	7.891	98	157854	3.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.800%
43) trans-1,3-Dichloroprop...	8.177	79	18347	3.397	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.000%
46) 2-Hexanone-d5	8.627	63	87591	39.594	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.180%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	43002	4.298	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	53100	4.399	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.000%
Target Compounds						
					Qvalue	
35) Methylcyclohexane	6.679	83	13097	0.626	ug/L #	21
37) 1,2-Dichloropropane	6.682	63	5988	0.489	ug/L #	89
61) 1,2,3-Trichloropropane	11.804	75	8427	1.241	ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061393.D
Acq On : 02 Nov 2024 04:31
Operator : MD/SY
Sample : P4668-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Nov 04 05:10:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

