

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061361.D
 Acq On : 01 Nov 2024 14:37
 Operator : MD/SY
 Sample : P4631-19
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP13

Quant Time: Nov 04 04:00:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	163497	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	148886	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	64834	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	40870	3.680	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.600%
7) Chloroethane-d5	1.907	69	42311	4.156	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.560	65	19099	4.019	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.400%
20) 2-Butanone-d5	4.624	46	108385	48.066	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.140%
24) Chloroform-d	5.055	84	97941	4.646	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.695	65	51293	4.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	5.721	84	196725	4.689	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
36) 1,2-Dichloropropane-d6	6.682	67	61436	4.871	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.400%
41) Toluene-d8	7.891	98	178666	4.555	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
43) trans-1,3-Dichloroprop...	8.174	79	21507	4.177	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.627	63	93282	44.237	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.480%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	45478	4.768	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	58159	5.066	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%
Target Compounds					Qvalue	
13) Acetone	2.644	43	9190	6.435	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061361.D
Acq On : 01 Nov 2024 14:37
Operator : MD/SY
Sample : P4631-19
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP13

Quant Time: Nov 04 04:00:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

