

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061350.D
 Acq On : 01 Nov 2024 09:55
 Operator : MD/SY
 Sample : VU1101WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK071

Quant Time: Nov 04 03:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	163453	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	151264	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	67264	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	53219	4.793	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.800%
7) Chloroethane-d5	1.911	69	51511	5.060	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.200%
11) 1,1-Dichloroethene-d2	2.560	65	23327	4.910	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.200%
20) 2-Butanone-d5	4.627	46	117573	52.155	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.300%
24) Chloroform-d	5.052	84	107759	5.113	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.200%
26) 1,2-Dichloroethane-d4	5.695	65	55894	5.321	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.400%
32) Benzene-d6	5.717	84	223850	5.252	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.000%
36) 1,2-Dichloropropane-d6	6.682	67	68658	5.358	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.200%
41) Toluene-d8	7.891	98	202596	5.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
43) trans-1,3-Dichloroprop...	8.177	79	23672	4.526	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.600%
46) 2-Hexanone-d5	8.627	63	103453	48.289	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.580%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	48131	4.967	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	61701	5.180	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061350.D
Acq On : 01 Nov 2024 09:55
Operator : MD/SY
Sample : VU1101WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK071

Quant Time: Nov 04 03:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

