

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
 Data File : VU061382.D
 Acq On : 02 Nov 2024 00:02
 Operator : MD/SY
 Sample : P4671-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AB6

Quant Time: Nov 04 05:08:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 01 00:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	168132	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	155459	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	68181	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	33555	2.938	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	58.800%
7) Chloroethane-d5	1.907	69	39499	3.772	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.400%
11) 1,1-Dichloroethene-d2	2.560	65	16505	3.378	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	4.624	46	98914	42.656	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.320%
24) Chloroform-d	5.055	84	85859	3.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
26) 1,2-Dichloroethane-d4	5.695	65	43887	4.061	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
32) Benzene-d6	5.717	84	173203	3.954	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.000%
36) 1,2-Dichloropropane-d6	6.682	67	54500	4.138	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.800%
41) Toluene-d8	7.891	98	154417	3.770	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.174	79	17607	3.275	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.600%
46) 2-Hexanone-d5	8.627	63	83443	37.898	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.800%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	41712	4.188	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	52412	4.341	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110124\
Data File : VU061382.D
Acq On : 02 Nov 2024 00:02
Operator : MD/SY
Sample : P4671-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AB6

Quant Time: Nov 04 05:08:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 01 00:44:02 2024
Response via : Initial Calibration

