

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110524\
 Data File : VU061452.D
 Acq On : 05 Nov 2024 14:57
 Operator : MD/SY
 Sample : VIBLK156
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK156

Quant Time: Nov 06 00:02:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 05 23:59:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	154120	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	145864	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63608	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	39620	3.784	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.600%
7) Chloroethane-d5	1.907	69	44690	4.656	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.200%
11) 1,1-Dichloroethene-d2	2.560	65	18655	4.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.200%
20) 2-Butanone-d5	4.624	46	109531	51.529	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.060%
24) Chloroform-d	5.052	84	95350	4.798	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	5.695	65	49601	5.008	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	5.717	84	192200	4.676	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
36) 1,2-Dichloropropane-d6	6.682	67	60703	4.913	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.891	98	170100	4.426	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
43) trans-1,3-Dichloroprop...	8.174	79	18944	3.756	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%
46) 2-Hexanone-d5	8.627	63	95630	46.290	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.580%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	45409	4.860	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	56780	5.041	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110524\
Data File : VU061452.D
Acq On : 05 Nov 2024 14:57
Operator : MD/SY
Sample : VIBLK156
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK156

Quant Time: Nov 06 00:02:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 05 23:59:28 2024
Response via : Initial Calibration

