

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110524\
 Data File : VU061453.D
 Acq On : 05 Nov 2024 15:22
 Operator : MD/SY
 Sample : P4671-05
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 06 00:03:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 05 23:59:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	126544	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	117328	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	51534	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	50988	5.932	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.600%
7) Chloroethane-d5	1.908	69	50272	6.379	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	127.600%
11) 1,1-Dichloroethene-d2	2.560	65	22367	6.082	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	121.600%
20) 2-Butanone-d5	4.624	46	101993	58.439	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.880%
24) Chloroform-d	5.055	84	98579	6.042	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	120.800%
26) 1,2-Dichloroethane-d4	5.692	65	50884	6.257	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.200%
32) Benzene-d6	5.718	84	207809	6.286	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	125.800%#
36) 1,2-Dichloropropane-d6	6.682	67	62024	6.240	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	124.800%
41) Toluene-d8	7.891	98	182920	5.918	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.400%
43) trans-1,3-Dichloroprop...	8.174	79	20237	4.988	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.800%
46) 2-Hexanone-d5	8.627	63	94106	56.632	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.260%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	45093	5.999	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	58426	6.402	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	128.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110524\
Data File : VU061453.D
Acq On : 05 Nov 2024 15:22
Operator : MD/SY
Sample : P4671-05
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 06 00:03:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 05 23:59:28 2024
Response via : Initial Calibration

