

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
 Data File : VU061494.D
 Acq On : 06 Nov 2024 12:02
 Operator : MD/SY
 Sample : P4714-11
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAH2

Quant Time: Nov 07 00:51:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 07 00:10:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	148644	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	138003	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	60304	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	47642	4.718	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.400%
7) Chloroethane-d5	1.908	69	50981	5.507	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.200%
11) 1,1-Dichloroethene-d2	2.560	65	22604	5.232	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.600%
20) 2-Butanone-d5	4.625	46	109286	53.308	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.620%
24) Chloroform-d	5.052	84	101196	5.280	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
26) 1,2-Dichloroethane-d4	5.695	65	53598	5.610	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.200%
32) Benzene-d6	5.718	84	212718	5.470	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.400%
36) 1,2-Dichloropropane-d6	6.682	67	64797	5.543	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.800%
41) Toluene-d8	7.891	98	186937	5.141	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.800%
43) trans-1,3-Dichloroprop...	8.174	79	20370	4.269	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.400%
46) 2-Hexanone-d5	8.628	63	98694	50.495	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.980%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	48567	5.494	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	60339	5.650	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.000%
Target Compounds						
13) Acetone	2.644	43	6346	4.887	ug/L	95
16) Methylene chloride	3.043	84	10724	1.036	ug/L	92
42) Toluene	7.968	91	8247	0.178	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
Data File : VU061494.D
Acq On : 06 Nov 2024 12:02
Operator : MD/SY
Sample : P4714-11
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH2

Quant Time: Nov 07 00:51:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 07 00:10:32 2024
Response via : Initial Calibration

