

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
 Data File : VU061495.D
 Acq On : 06 Nov 2024 12:27
 Operator : MD/SY
 Sample : VIBLK157
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK157

Quant Time: Nov 07 00:51:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 07 00:10:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	169424	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	159377	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	71547	5.000	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.595	65	39342	3.418	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.400%
7) Chloroethane-d5	1.911	69	42847	4.061	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.200%
11) 1,1-Dichloroethene-d2	2.560	65	19225	3.904	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.000%
20) 2-Butanone-d5	4.627	46	111052	47.526	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.060%
24) Chloroform-d	5.052	84	97616	4.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.695	65	50647	4.651	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
32) Benzene-d6	5.717	84	197592	4.400	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
36) 1,2-Dichloropropane-d6	6.682	67	62748	4.648	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.891	98	173609	4.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	8.174	79	19865	3.604	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.000%
46) 2-Hexanone-d5	8.627	63	100689	44.607	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.220%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	47477	4.650	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	60198	4.751	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
Data File : VU061495.D
Acq On : 06 Nov 2024 12:27
Operator : MD/SY
Sample : VIBLK157
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK157

Quant Time: Nov 07 00:51:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 07 00:10:32 2024
Response via : Initial Calibration

