

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
 Data File : VU061496.D
 Acq On : 06 Nov 2024 12:52
 Operator : MD/SY
 Sample : P4689-08
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CD2

Quant Time: Nov 07 00:51:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 07 00:10:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	152817	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	144087	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	61562	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	45805	4.412	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.200%
7) Chloroethane-d5	1.907	69	50013	5.255	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.560	65	21363	4.810	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.200%
20) 2-Butanone-d5	4.624	46	109971	52.178	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.360%
24) Chloroform-d	5.055	84	98313	4.989	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.695	65	50971	5.190	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	5.717	84	206473	5.086	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
36) 1,2-Dichloropropane-d6	6.682	67	61870	5.069	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.891	98	180575	4.757	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	8.174	79	20196	4.053	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.627	63	97974	48.010	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.020%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	47614	5.158	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	56811	5.211	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
Data File : VU061496.D
Acq On : 06 Nov 2024 12:52
Operator : MD/SY
Sample : P4689-08
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CD2

Quant Time: Nov 07 00:51:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 07 00:10:32 2024
Response via : Initial Calibration

