

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
 Data File : VU061500.D
 Acq On : 06 Nov 2024 14:30
 Operator : MD/SY
 Sample : P4625-02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AA2

Quant Time: Nov 07 00:52:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 07 00:10:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	145535	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	139031	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	59071	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	47385	4.793	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.800%
7) Chloroethane-d5	1.907	69	50225	5.542	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.800%
11) 1,1-Dichloroethene-d2	2.560	65	21630	5.114	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.200%
20) 2-Butanone-d5	4.624	46	107864	53.739	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.480%
24) Chloroform-d	5.052	84	100728	5.368	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.400%
26) 1,2-Dichloroethane-d4	5.692	65	52062	5.566	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.400%
32) Benzene-d6	5.721	84	210530	5.374	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.400%
36) 1,2-Dichloropropane-d6	6.682	67	63836	5.420	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.400%
41) Toluene-d8	7.891	98	186961	5.104	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%
43) trans-1,3-Dichloroprop...	8.177	79	20045	4.169	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.627	63	98273	49.908	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.820%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	47765	5.363	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	62106	5.937	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
Data File : VU061500.D
Acq On : 06 Nov 2024 14:30
Operator : MD/SY
Sample : P4625-02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AA2

Quant Time: Nov 07 00:52:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 07 00:10:32 2024
Response via : Initial Calibration

