

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
 Data File : VU061509.D
 Acq On : 06 Nov 2024 18:11
 Operator : MD/SY
 Sample : P4714-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AG4

Quant Time: Nov 07 00:55:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 07 00:10:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	138643	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	130931	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	56028	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	44936	4.771	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.400%
7) Chloroethane-d5	1.907	69	46221	5.353	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.000%
11) 1,1-Dichloroethene-d2	2.560	65	20466	5.079	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.600%
20) 2-Butanone-d5	4.624	46	104021	54.400	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.800%
24) Chloroform-d	5.055	84	97130	5.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.600%
26) 1,2-Dichloroethane-d4	5.695	65	51088	5.733	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.600%
32) Benzene-d6	5.717	84	200871	5.445	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.800%
36) 1,2-Dichloropropane-d6	6.682	67	60856	5.487	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.800%
41) Toluene-d8	7.891	98	178805	5.183	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
43) trans-1,3-Dichloroprop...	8.177	79	18466	4.079	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.627	63	94087	50.738	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.480%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	45964	5.480	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	58593	5.906	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110624\
Data File : VU061509.D
Acq On : 06 Nov 2024 18:11
Operator : MD/SY
Sample : P4714-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AG4

Quant Time: Nov 07 00:55:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 07 00:10:32 2024
Response via : Initial Calibration

