

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045606.D
 Acq On : 08 Nov 2021 11:42
 Operator : SY/MD
 Sample : M4434-13 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 2004

Quant Time: Nov 08 14:58:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	131264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	246497	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	243284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	119189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	108809	55.350	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.700%	
35) Dibromofluoromethane	5.292	113	83512	50.745	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.480%	
50) Toluene-d8	7.899	98	323395	52.846	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.700%	
62) 4-Bromofluorobenzene	10.636	95	119601	48.738	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.480%	
Target Compounds						Qvalue
16) Acetone	2.626	43	10670	10.662	ug/l	100
89) n-Butylbenzene	12.211	91	1521	0.220	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045606.D
Acq On : 08 Nov 2021 11:42
Operator : SY/MD
Sample : M4434-13 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
2004

Quant Time: Nov 08 14:58:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

