

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045613.D
 Acq On : 08 Nov 2021 14:23
 Operator : SY/MD
 Sample : M4571-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-4

Quant Time: Nov 09 09:56:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	162755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	298101	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	305389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	159285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	127989	52.510	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.020%	
35) Dibromofluoromethane	5.292	113	100736	50.614	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.220%	
50) Toluene-d8	7.899	98	391568	52.909	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.820%	
62) 4-Bromofluorobenzene	10.636	95	153461	51.711	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.420%	
Target Compounds						
64) Tetrachloroethene	8.558	164	4939	2.818	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045613.D
Acq On : 08 Nov 2021 14:23
Operator : SY/MD
Sample : M4571-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-4

Quant Time: Nov 09 09:56:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

