

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045615.D
 Acq On : 08 Nov 2021 15:09
 Operator : SY/MD
 Sample : M4571-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-5

Quant Time: Nov 09 09:57:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	160510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	302759	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	304961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	159669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	129903	54.040	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.080%	
35) Dibromofluoromethane	5.292	113	101743	50.334	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.660%	
50) Toluene-d8	7.899	98	394382	52.469	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.940%	
62) 4-Bromofluorobenzene	10.635	95	155102	51.460	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.920%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.671	96	6769	3.137	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045615.D
Acq On : 08 Nov 2021 15:09
Operator : SY/MD
Sample : M4571-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-5

Quant Time: Nov 09 09:57:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

