

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045616.D
 Acq On : 08 Nov 2021 15:32
 Operator : SY/MD
 Sample : M4571-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-1

Quant Time: Nov 09 09:57:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	158203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	294508	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	297227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	148042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	125600	53.012	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.020%	
35) Dibromofluoromethane	5.292	113	99442	50.574	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.140%	
50) Toluene-d8	7.899	98	386363	52.843	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.680%	
62) 4-Bromofluorobenzene	10.635	95	148970	50.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.620%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.671	96	82904	38.981	ug/l	89
44) Trichloroethene	6.546	130	7822	3.844	ug/l	85
64) Tetrachloroethene	8.555	164	18025	10.566	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045616.D
Acq On : 08 Nov 2021 15:32
Operator : SY/MD
Sample : M4571-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-1

Quant Time: Nov 09 09:57:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

