

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110821\  
 Data File : VU045618.D  
 Acq On : 08 Nov 2021 16:18  
 Operator : SY/MD  
 Sample : M4523-08  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 TRIP-BLANK

Quant Time: Nov 09 09:58:46 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U110521W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 05 13:35:16 2021  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|-----------|-------|----------|----------|-------|-----------|
| -----                       |           |       |          |          |       |           |
| Internal Standards          |           |       |          |          |       |           |
| 1) Pentafluorobenzene       | 5.379     | 168   | 159274   | 50.000   | ug/l  | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.253     | 114   | 298688   | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 9.420     | 117   | 301891   | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 11.816    | 152   | 156666   | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |           |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 5.703     | 65    | 126401   | 52.991   | ug/l  | 0.00      |
| Spiked Amount               | 50.000    | Range | 78 - 117 | Recovery | =     | 105.980%  |
| 35) Dibromofluoromethane    | 5.292     | 113   | 98773    | 49.531   | ug/l  | 0.00      |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | =     | 99.060%   |
| 50) Toluene-d8              | 7.899     | 98    | 390534   | 52.666   | ug/l  | 0.00      |
| Spiked Amount               | 50.000    | Range | 92 - 112 | Recovery | =     | 105.340%  |
| 62) 4-Bromofluorobenzene    | 10.636    | 95    | 152582   | 51.314   | ug/l  | 0.00      |
| Spiked Amount               | 50.000    | Range | 83 - 123 | Recovery | =     | 102.620%  |
| Target Compounds            |           |       |          |          |       |           |
| 16) Acetone                 | 2.626     | 43    | 2345     | 1.931    | ug/l  | Qvalue 99 |
| -----                       |           |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110821\  
Data File : VU045618.D  
Acq On : 08 Nov 2021 16:18  
Operator : SY/MD  
Sample : M4523-08  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
TRIP-BLANK

Quant Time: Nov 09 09:58:46 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U110521W.M  
Quant Title : SW846 8260  
QLast Update : Fri Nov 05 13:35:16 2021  
Response via : Initial Calibration

