

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045619.D
 Acq On : 08 Nov 2021 16:41
 Operator : SY/MD
 Sample : M4523-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FIELD-BLANK

Quant Time: Nov 09 09:59:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	156575	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	298570	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	297193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	155820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	125706	53.609	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 107.220%			
35) Dibromofluoromethane	5.292	113	98550	49.438	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.880%			
50) Toluene-d8	7.899	98	384298	51.845	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.700%			
62) 4-Bromofluorobenzene	10.635	95	151540	50.983	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.960%			
Target Compounds					Qvalue	
16) Acetone	2.629	43	4150	3.477	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045619.D
Acq On : 08 Nov 2021 16:41
Operator : SY/MD
Sample : M4523-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FIELD-BLANK

Quant Time: Nov 09 09:59:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

