

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110821\  
 Data File : VU045620.D  
 Acq On : 08 Nov 2021 17:04  
 Operator : SY/MD  
 Sample : M4504-05  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 RW8-MW01D2-20211103

Quant Time: Nov 09 09:59:25 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U110521W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 05 13:35:16 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min)      |
|-----------------------------|--------|-------|----------|----------|------------|---------------|
| -----                       |        |       |          |          |            |               |
| Internal Standards          |        |       |          |          |            |               |
| 1) Pentafluorobenzene       | 5.379  | 168   | 155227   | 50.000   | ug/l       | 0.00          |
| 34) 1,4-Difluorobenzene     | 6.253  | 114   | 291498   | 50.000   | ug/l       | 0.00          |
| 63) Chlorobenzene-d5        | 9.420  | 117   | 295528   | 50.000   | ug/l       | 0.00          |
| 72) 1,4-Dichlorobenzene-d4  | 11.812 | 152   | 150831   | 50.000   | ug/l       | 0.00          |
| System Monitoring Compounds |        |       |          |          |            |               |
| 33) 1,2-Dichloroethane-d4   | 5.703  | 65    | 124451   | 53.534   | ug/l       | 0.00          |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 107.060% |               |
| 35) Dibromofluoromethane    | 5.292  | 113   | 98083    | 50.398   | ug/l       | 0.00          |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 100.800% |               |
| 50) Toluene-d8              | 7.899  | 98    | 381092   | 52.660   | ug/l       | 0.00          |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 105.320% |               |
| 62) 4-Bromofluorobenzene    | 10.632 | 95    | 148857   | 51.296   | ug/l       | 0.00          |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 102.600% |               |
| Target Compounds            |        |       |          |          |            |               |
| 16) Acetone                 | 2.626  | 43    | 4038     | 3.412    | ug/l       | Qvalue<br>100 |
| -----                       |        |       |          |          |            |               |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110821\  
Data File : VU045620.D  
Acq On : 08 Nov 2021 17:04  
Operator : SY/MD  
Sample : M4504-05  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
RW8-MW01D2-20211103

Quant Time: Nov 09 09:59:25 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U110521W.M  
Quant Title : SW846 8260  
QLast Update : Fri Nov 05 13:35:16 2021  
Response via : Initial Calibration

