

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045623.D
 Acq On : 08 Nov 2021 18:13
 Operator : SY/MD
 Sample : M4485-11
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 AG-7

Quant Time: Nov 09 10:00:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	147825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	277197	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	283311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	146784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	119849	54.136	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.280%	
35) Dibromofluoromethane	5.292	113	93401	50.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.940%	
50) Toluene-d8	7.899	98	363183	52.774	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.540%	
62) 4-Bromofluorobenzene	10.636	95	143871	52.135	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.280%	
Target Compounds						
16) Acetone	2.626	43	5800	5.147	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045623.D
Acq On : 08 Nov 2021 18:13
Operator : SY/MD
Sample : M4485-11
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
AG-7

Quant Time: Nov 09 10:00:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

