

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045624.D
 Acq On : 08 Nov 2021 18:36
 Operator : SY/MD
 Sample : M4485-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 AG-9

Quant Time: Nov 09 10:01:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	148338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	282976	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	287211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	147401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	123523	55.603	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.200%
35) Dibromofluoromethane	5.292	113	94441	49.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	7.903	98	368048	52.389	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.780%
62) 4-Bromofluorobenzene	10.636	95	142288	50.509	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.020%
Target Compounds						
16) Acetone	2.626	43	3527	3.119	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045624.D
Acq On : 08 Nov 2021 18:36
Operator : SY/MD
Sample : M4485-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
AG-9

Quant Time: Nov 09 10:01:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

