

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045632.D
 Acq On : 08 Nov 2021 22:27
 Operator : SY/MD
 Sample : VU1108WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1108WBL02

Quant Time: Nov 09 10:06:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	153460	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	283233	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	287092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	145493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	120864	52.590	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.180%	
35) Dibromofluoromethane	5.292	113	94867	50.168	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.340%	
50) Toluene-d8	7.900	98	371539	52.838	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.680%	
62) 4-Bromofluorobenzene	10.636	95	143056	50.735	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045632.D
Acq On : 08 Nov 2021 22:27
Operator : SY/MD
Sample : VU1108WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1108WBL02

Quant Time: Nov 09 10:06:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

