

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045638.D
 Acq On : 09 Nov 2021 00:45
 Operator : SY/MD
 Sample : M4495-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DRW-2

Quant Time: Nov 09 11:15:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	151489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	286122	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	292207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	151628	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	122379	53.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.880%
35) Dibromofluoromethane	5.292	113	96576	50.556	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.120%
50) Toluene-d8	7.899	98	376355	52.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.960%
62) 4-Bromofluorobenzene	10.636	95	148134	52.006	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045638.D
Acq On : 09 Nov 2021 00:45
Operator : SY/MD
Sample : M4495-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DRW-2

Quant Time: Nov 09 11:15:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

