

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045640.D
 Acq On : 09 Nov 2021 01:31
 Operator : SY/MD
 Sample : M4504-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RW8-MW01S-20211102

Quant Time: Nov 09 11:15:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	149065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	281858	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	280284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	140960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	119504	53.531	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.060%	
35) Dibromofluoromethane	5.292	113	93060	49.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.900%	
50) Toluene-d8	7.899	98	365048	52.168	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.340%	
62) 4-Bromofluorobenzene	10.635	95	139900	49.858	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.720%	
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	5195	4.571	ug/l	89
29) Tetrahydrofuran	5.054	42	10316	10.700	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045640.D
Acq On : 09 Nov 2021 01:31
Operator : SY/MD
Sample : M4504-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RW8-MW01S-20211102

Quant Time: Nov 09 11:15:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

