

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110821\  
 Data File : VU045641.D  
 Acq On : 09 Nov 2021 01:54  
 Operator : SY/MD  
 Sample : M4504-03  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 44 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 RW8-MW01D1-20211102

Quant Time: Nov 09 11:16:06 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U110521W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 05 13:35:16 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.378  | 168   | 148970   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.253  | 114   | 279745   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 9.420  | 117   | 285659   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.812 | 152   | 146029   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.703  | 65    | 120534   | 54.027   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 108.060% |          |
| 35) Dibromofluoromethane    | 5.292  | 113   | 93252    | 49.929   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 99.860%  |          |
| 50) Toluene-d8              | 7.899  | 98    | 363351   | 52.318   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 104.640% |          |
| 62) 4-Bromofluorobenzene    | 10.636 | 95    | 140942   | 50.609   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 101.220% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110821\  
Data File : VU045641.D  
Acq On : 09 Nov 2021 01:54  
Operator : SY/MD  
Sample : M4504-03  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 44 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
RW8-MW01D1-20211102

Quant Time: Nov 09 11:16:06 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U110521W.M  
Quant Title : SW846 8260  
QLast Update : Fri Nov 05 13:35:16 2021  
Response via : Initial Calibration

