

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045643.D
 Acq On : 09 Nov 2021 02:40
 Operator : SY/MD
 Sample : M4504-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP01-20211103

Quant Time: Nov 09 11:16:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	147536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	279206	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	281776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	142512	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	118800	53.767	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.540%
35) Dibromofluoromethane	5.292	113	92054	49.382	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.760%
50) Toluene-d8	7.899	98	361855	52.203	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.400%
62) 4-Bromofluorobenzene	10.635	95	140846	50.672	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045643.D
Acq On : 09 Nov 2021 02:40
Operator : SY/MD
Sample : M4504-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP01-20211103

Quant Time: Nov 09 11:16:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

