

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045644.D
 Acq On : 09 Nov 2021 03:03
 Operator : SY/MD
 Sample : PB140545TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140545TB

Quant Time: Nov 09 11:17:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	147367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	282713	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	288865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	143674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	119883	54.320	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.640%	
35) Dibromofluoromethane	5.292	113	94284	49.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.900%	
50) Toluene-d8	7.899	98	374037	53.291	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.580%	
62) 4-Bromofluorobenzene	10.636	95	143774	51.084	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.160%	
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	15633	13.915	ug/l	97
20) Methylene Chloride	3.051	84	7330	3.755	ug/l	98
43) Isopropyl Acetate	5.912	43	21118	4.360	ug/l	99
80) 1,3,5-Trimethylbenzene	11.092	105	4729	0.577	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045644.D
Acq On : 09 Nov 2021 03:03
Operator : SY/MD
Sample : PB140545TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140545TB

Quant Time: Nov 09 11:17:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

