

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
 Data File : VU045645.D
 Acq On : 09 Nov 2021 03:26
 Operator : SY/MD
 Sample : M4541-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-3-AS

Quant Time: Nov 09 11:17:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	145659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	272575	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	281489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	140532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	117430	53.832	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.660%	
35) Dibromofluoromethane	5.292	113	91646	50.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.720%	
50) Toluene-d8	7.899	98	363545	53.723	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.440%	
62) 4-Bromofluorobenzene	10.636	95	139102	51.262	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.520%	
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	16568	14.920	ug/l	95
20) Methylene Chloride	3.051	84	7979	4.135	ug/l	96
40) Benzene	5.774	78	3446	0.438	ug/l	99
43) Isopropyl Acetate	5.912	43	19528	4.182	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110821\
Data File : VU045645.D
Acq On : 09 Nov 2021 03:26
Operator : SY/MD
Sample : M4541-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-3-AS

Quant Time: Nov 09 11:17:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

