

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
 Data File : VU061558.D
 Acq On : 08 Nov 2024 19:26
 Operator : MD/SY
 Sample : P4748-05
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP59

Quant Time: Nov 09 00:23:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 00:17:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	159697	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	151304	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	64814	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	36436	3.359	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.200%
7) Chloroethane-d5	1.911	69	39184	3.940	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.563	65	17366	3.742	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.800%
20) 2-Butanone-d5	4.627	46	99859	45.339	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.680%
24) Chloroform-d	5.055	84	85532	4.154	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
26) 1,2-Dichloroethane-d4	5.695	65	44744	4.359	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
32) Benzene-d6	5.717	84	172745	4.052	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.000%
36) 1,2-Dichloropropane-d6	6.682	67	54873	4.281	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.600%
41) Toluene-d8	7.891	98	149639	3.754	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.000%
43) trans-1,3-Dichloroprop...	8.174	79	15938	3.046	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.000%
46) 2-Hexanone-d5	8.627	63	83992	39.195	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.400%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	41138	4.244	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	49755	4.335	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.800%
Target Compounds					Qvalue	
13) Acetone	2.644	43	13006	9.323	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
Data File : VU061558.D
Acq On : 08 Nov 2024 19:26
Operator : MD/SY
Sample : P4748-05
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP59

Quant Time: Nov 09 00:23:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 00:17:10 2024
Response via : Initial Calibration

