

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
 Data File : VU061563.D
 Acq On : 08 Nov 2024 22:17
 Operator : MD/SY
 Sample : VU1108WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK080

Quant Time: Nov 09 01:40:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 01:40:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	158543	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	147060	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	65574	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	33909	3.149	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.000%
7) Chloroethane-d5	1.911	69	36370	3.684	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.600%
11) 1,1-Dichloroethene-d2	2.560	65	16410	3.561	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.200%
20) 2-Butanone-d5	4.631	46	113983	52.128	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.260%
24) Chloroform-d	5.052	84	94017	4.599	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	5.692	65	48104	4.721	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	5.718	84	184247	4.446	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
36) 1,2-Dichloropropane-d6	6.682	67	58406	4.688	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.891	98	156617	4.042	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	8.174	79	17549	3.451	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.000%
46) 2-Hexanone-d5	8.627	63	100034	48.028	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.060%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	46728	4.960	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	54419	4.687	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
Data File : VU061563.D
Acq On : 08 Nov 2024 22:17
Operator : MD/SY
Sample : VU1108WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK080

Quant Time: Nov 09 01:40:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 01:40:03 2024
Response via : Initial Calibration

