

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
 Data File : VU061570.D
 Acq On : 09 Nov 2024 01:22
 Operator : MD/SY
 Sample : P4748-13
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP67

Quant Time: Nov 11 00:10:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 01:40:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	165598	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	156004	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	67726	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	39421	3.504	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.000%
7) Chloroethane-d5	1.911	69	41354	4.010	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.200%
11) 1,1-Dichloroethene-d2	2.560	65	18734	3.892	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.800%
20) 2-Butanone-d5	4.624	46	106038	46.428	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.860%
24) Chloroform-d	5.052	84	93527	4.380	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
26) 1,2-Dichloroethane-d4	5.695	65	47071	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
32) Benzene-d6	5.718	84	188046	4.278	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.682	67	59048	4.468	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.400%
41) Toluene-d8	7.891	98	165354	4.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	8.174	79	18785	3.482	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.600%
46) 2-Hexanone-d5	8.628	63	91906	41.596	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.200%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	45908	4.594	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	54904	4.578	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.600%
Target Compounds						
13) Acetone	2.644	43	14446	9.986	ug/L	92
25) Chloroform	5.078	83	14815	0.690	ug/L	97
31) Carbon tetrachloride	5.515	117	967412	59.183	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
Data File : VU061570.D
Acq On : 09 Nov 2024 01:22
Operator : MD/SY
Sample : P4748-13
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP67

Quant Time: Nov 11 00:10:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 01:40:03 2024
Response via : Initial Calibration

