

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
 Data File : VU061575.D
 Acq On : 09 Nov 2024 03:24
 Operator : MD/SY
 Sample : P4748-18
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP72

Quant Time: Nov 11 00:12:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 01:40:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	167344	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	159057	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	70242	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	39946	3.514	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.200%
7) Chloroethane-d5	1.911	69	37042	3.554	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.000%
11) 1,1-Dichloroethene-d2	2.557	65	22671	4.661	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.200%
20) 2-Butanone-d5	4.650	46	90008	38.999	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.000%
24) Chloroform-d	5.052	84	91814	4.255	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	5.695	65	46901	4.361	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
32) Benzene-d6	5.717	84	191192	4.266	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.400%
36) 1,2-Dichloropropane-d6	6.682	67	58943	4.375	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.891	98	166623	3.976	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	8.174	79	17105	3.110	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.200%
46) 2-Hexanone-d5	8.631	63	92529	41.074	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.140%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	45241	4.440	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	57680	4.637	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%
Target Compounds					Qvalue	
13) Acetone	2.673	43	267404	182.926	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
Data File : VU061575.D
Acq On : 09 Nov 2024 03:24
Operator : MD/SY
Sample : P4748-18
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP72

Quant Time: Nov 11 00:12:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 01:40:03 2024
Response via : Initial Calibration

