

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
 Data File : VU061554.D
 Acq On : 08 Nov 2024 17:48
 Operator : MD/SY
 Sample : P4748-01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP55

Quant Time: Nov 09 00:21:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 00:17:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	173599	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	165355	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	71275	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	41569	3.525	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.600%
7) Chloroethane-d5	1.911	69	45105	4.172	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.400%
11) 1,1-Dichloroethene-d2	2.560	65	19712	3.907	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.200%
20) 2-Butanone-d5	4.624	46	114446	47.800	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.600%
24) Chloroform-d	5.055	84	95118	4.249	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	5.695	65	50236	4.503	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
32) Benzene-d6	5.721	84	198046	4.251	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
36) 1,2-Dichloropropane-d6	6.682	67	60262	4.302	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.000%
41) Toluene-d8	7.891	98	171615	3.939	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	8.177	79	18053	3.157	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.200%
46) 2-Hexanone-d5	8.627	63	98063	41.873	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.740%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	47667	4.500	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	56965	4.513	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%
Target Compounds					Qvalue	
13) Acetone	2.644	43	14922	9.840	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110824\
Data File : VU061554.D
Acq On : 08 Nov 2024 17:48
Operator : MD/SY
Sample : P4748-01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCP55


Quant Time: Nov 09 00:21:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 00:17:10 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 9.840 ug/L
 RT: 2.644 min Scan# 42
 Delta R.T. 0.003 min
 Lab File: VU061554.D
 Acq: 08 Nov 2024 17:48

Instrument :
 MSVOA_U
 ClientSampleId :
 GCP55



Tgt Ion: 43 Resp: 14922
 Ion Ratio Lower Upper
 43 100
 58 36.9 0.0 73.2

