

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111123\
 Data File : VU056220.D
 Acq On : 11 Nov 2023 16:28
 Operator : MD/SY
 Sample : 05339-15RE
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCDK0RE

Quant Time: Nov 11 20:23:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 11 20:19:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	353690	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	351523	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	158357	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	44529	2.028	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	40.600%
7) Chloroethane-d5	1.907	69	38343	2.574	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	51.400%#
11) 1,1-Dichloroethene-d2	2.563	65	21308	2.203	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	44.000%#
20) 2-Butanone-d5	4.621	46	123704	34.347	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	68.700%
24) Chloroform-d	5.058	84	110908	2.654	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	53.000%#
26) 1,2-Dichloroethane-d4	5.698	65	53096	2.886	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	57.800%#
32) Benzene-d6	5.721	84	209374	2.480	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	49.600%#
36) 1,2-Dichloropropane-d6	6.685	67	66897	2.692	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	53.800%#
41) Toluene-d8	7.894	98	174517	2.256	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	45.200%#
43) trans-1,3-Dichloroprop...	8.177	79	23578	2.482	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	49.600%#
46) 2-Hexanone-d5	8.631	63	98364	31.591	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.180%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	53457	3.014	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	60.200%#
66) 1,2-Dichlorobenzene-d4	12.190	152	65137	2.667	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	53.400%#
Target Compounds						
13) Acetone	2.628	43	83327	31.371	ug/L	97
31) Carbon tetrachloride	5.518	117	40052	1.120	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111123\
Data File : VU056220.D
Acq On : 11 Nov 2023 16:28
Operator : MD/SY
Sample : 05339-15RE
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCDK0RE

Quant Time: Nov 11 20:23:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 11 20:19:10 2023
Response via : Initial Calibration

