

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111123\
 Data File : VU056230.D
 Acq On : 11 Nov 2023 21:19
 Operator : MD/SY
 Sample : 05371-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCDN1

Quant Time: Nov 11 22:19:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 11 22:15:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	367670	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	359460	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	164336	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	86966	3.811	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.200%
7) Chloroethane-d5	1.910	69	76905	4.966	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.400%
11) 1,1-Dichloroethene-d2	2.563	65	41723	4.150	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.000%
20) 2-Butanone-d5	4.627	46	276850	73.946	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	147.900%#
24) Chloroform-d	5.058	84	217538	5.008	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.695	65	104190	5.447	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.000%
32) Benzene-d6	5.724	84	423901	4.910	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	6.685	67	133202	5.241	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.800%
41) Toluene-d8	7.894	98	350643	4.432	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
43) trans-1,3-Dichloroprop...	8.177	79	46611	4.798	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.000%
46) 2-Hexanone-d5	8.627	63	239220	75.133	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	150.260%#
56) 1,1,2,2-Tetrachloroeth...	10.753	84	108673	5.991	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	119.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	138580	5.468	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.400%
Target Compounds						Qvalue
3) Chloromethane	1.518	50	11619	0.398	ug/L	100
8) Chloroethane	1.930	64	6242	0.381	ug/L	90
13) Acetone	2.637	43	210099	76.092	ug/L	100
17) Methyl tert-butyl Ether	3.357	73	16008	0.301	ug/L	96
21) 2-Butanone	4.714	43	62604	13.748	ug/L	95
22) cis-1,2-Dichloroethene	4.660	96	11437	0.423	ug/L	97
34) Trichloroethene	6.537	95	104178	3.706	ug/L	95
47) Tetrachloroethene	8.547	164	882661	41.831	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111123\
Data File : VU056230.D
Acq On : 11 Nov 2023 21:19
Operator : MD/SY
Sample : 05371-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCDN1

Quant Time: Nov 11 22:19:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 11 22:15:12 2023
Response via : Initial Calibration

