

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111224\
 Data File : VU061603.D
 Acq On : 12 Nov 2024 14:44
 Operator : MD/SY
 Sample : P4635-01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WS

Quant Time: Nov 13 05:07:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U111224DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Nov 13 05:00:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.110	96	44472	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	13989	0.907	ug/l	0.00
Spiked Amount	1.000		Recovery	=	91.000%	
68) 1,2-Dichlorobenzene-d4	12.190	152	13735	0.894	ug/l	0.00
Spiked Amount	1.000		Recovery	=	89.000%	
Target Compounds						
						Qvalue
27) Chloroform	5.078	83	665759	30.690	ug/l	100
44) Bromodichloromethane	7.097	83	443486	30.771	ug/l	99
55) Dibromochloromethane	8.801	129	467784	49.396	ug/l	100
67) Bromoform	10.283	173	247137	50.305	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111224\
Data File : VU061603.D
Acq On : 12 Nov 2024 14:44
Operator : MD/SY
Sample : P4635-01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WS

Quant Time: Nov 13 05:07:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U111224DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Nov 13 05:00:10 2024
Response via : Initial Calibration

