

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111224\
 Data File : VU061610.D
 Acq On : 12 Nov 2024 18:38
 Operator : MD/SY
 Sample : P4635-01DL 3X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Nov 13 05:10:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U111224DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Nov 13 05:00:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Fluorobenzene	6.110	96	44458	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.630	95	11838	0.767	ug/l	0.00
Spiked Amount	1.000		Recovery	=	77.000%	
68) 1,2-Dichlorobenzene-d4	12.190	152	13072	0.851	ug/l	0.00
Spiked Amount	1.000		Recovery	=	85.000%	
Target Compounds						
						Qvalue
27) Chloroform	5.078	83	177853	8.201	ug/l	100
44) Bromodichloromethane	7.097	83	111655	7.749	ug/l	100
55) Dibromochloromethane	8.801	129	116028	12.256	ug/l	100
67) Bromoform	10.283	173	60120	12.241	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111224\
Data File : VU061610.D
Acq On : 12 Nov 2024 18:38
Operator : MD/SY
Sample : P4635-01DL 3X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Nov 13 05:10:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U111224DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Nov 13 05:00:10 2024
Response via : Initial Calibration









#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.110 min Scan# 1498
Delta R.T. 0.003 min
Lab File: VU061610.D
Acq: 12 Nov 2024 18:38




Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL


Tgt Ion:	96	Resp:	44458
Ion	Ratio	Lower	Upper
96	100		
70	18.5	15.8	23.6
95	8.6	7.3	10.9
97	6.8	0.0	0.0



#27
Chloroform
Concen: 8.201 ug/l
RT: 5.078 min Scan# 1178
Delta R.T. 0.003 min
Lab File: VU061610.D
Acq: 12 Nov 2024 18:38

Tgt Ion:	83	Resp:	177853
Ion	Ratio	Lower	Upper
83	100		
85	65.3	0.0	130.6









#44
 Bromodichloromethane
 Concen: 7.749 ug/l
 RT: 7.097 min Scan# 1806
 Delta R.T. 0.003 min
 Lab File: VU061610.D
 Acq: 12 Nov 2024 18:38




Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL


Tgt Ion:	83	Resp:	111655
Ion Ratio	Lower	Upper	
83	100		
85	64.0	51.4	77.0
127	9.4	7.2	10.8



#55
 Dibromochloromethane
 Concen: 12.256 ug/l
 RT: 8.801 min Scan# 2336
 Delta R.T. -0.000 min
 Lab File: VU061610.D
 Acq: 12 Nov 2024 18:38

Tgt Ion:	129	Resp:	116028
Ion Ratio	Lower	Upper	
129	100		
127	77.4	61.8	92.8





#57

4-Bromofluorobenzene

Concen: 0.767 ug/l

RT: 10.630 min Scan# 2905

Delta R.T. 0.006 min

Lab File: VU061610.D


Acq: 12 Nov 2024 18:38

Instrument :




MSVOA_U

ClientSampleId :

PT-THM-WSDL



Tgt Ion:	95	Resp:	11838
Ion Ratio	Lower	Upper	
95	100		
174	84.9	67.5	101.3
176	82.6	62.7	94.1



#67

Bromoform

Concen: 12.241 ug/l




RT: 10.283 min Scan# 2797


Delta R.T. 0.003 min

Lab File: VU061610.D

Acq: 12 Nov 2024 18:38


Tgt Ion:	173	Resp:	60120
Ion Ratio	Lower	Upper	
173	100		
175	47.9	38.4	57.6
254	9.0	7.7	11.5





#68
 1,2-Dichlorobenzene-d4
 Concen: 0.851 ug/l
 RT: 12.190 min Scan# 33
 Delta R.T. 0.006 min
 Lab File: VU061610.D
 Acq: 12 Nov 2024 18:38

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL



Tgt Ion:	152	Resp:	13072
Ion Ratio	Lower	Upper	
152	100		
115	63.2	0.0	259.0
150	153.2	0.0	650.2

