

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU111324\  
 Data File : VU061628.D  
 Acq On : 13 Nov 2024 16:52  
 Operator : MD/SY  
 Sample : P4747-06  
 Misc : 25mL/MSVOA\_U/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 GCPC5

Quant Time: Nov 14 01:31:22 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR111324WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Nov 14 00:27:20 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	168944	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	160293	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	69820	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	38154	3.495	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.000%	
7) Chloroethane-d5	1.907	69	43375	4.655	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.000%	
11) 1,1-Dichloroethene-d2	2.560	65	18726	3.824	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.400%	
20) 2-Butanone-d5	4.624	46	106086	45.788	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.580%	
24) Chloroform-d	5.052	84	94421	4.352	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.000%	
26) 1,2-Dichloroethane-d4	5.692	65	47903	4.601	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.000%	
32) Benzene-d6	5.718	84	192339	4.337	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6	6.682	67	60465	4.623	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.400%	
41) Toluene-d8	7.891	98	170172	4.117	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.400%	
43) trans-1,3-Dichloroprop...	8.174	79	16824	3.492	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	69.800%	
46) 2-Hexanone-d5	8.627	63	89874	47.210	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.420%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	45096	4.374	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	87.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	56641	4.555	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.000%	
Target Compounds						
13) Acetone	2.641	43	13421	9.194	ug/L	100
34) Trichloroethene	6.534	95	52214	3.986	ug/L	99
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU111324\  
 Data File : VU061628.D  
 Acq On : 13 Nov 2024 16:52  
 Operator : MD/SY  
 Sample : P4747-06  
 Misc : 25mL/MSVOA\_U/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 GCPC5

Quant Time: Nov 14 01:31:22 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR111324WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Nov 14 00:27:20 2024  
 Response via : Initial Calibration

