

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
 Data File : VU061630.D
 Acq On : 13 Nov 2024 17:41
 Operator : MD/SY
 Sample : P4747-08
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCPC7

Quant Time: Nov 14 01:31:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 14 00:27:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	172269	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	162420	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	70215	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	37586	3.377	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.600%
7) Chloroethane-d5	1.911	69	42173	4.438	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.800%
11) 1,1-Dichloroethene-d2	2.560	65	16909	3.386	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.800%
20) 2-Butanone-d5	4.624	46	110258	46.671	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.340%
24) Chloroform-d	5.052	84	95220	4.304	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	5.692	65	48005	4.521	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
32) Benzene-d6	5.718	84	188568	4.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	6.682	67	59463	4.487	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.800%
41) Toluene-d8	7.891	98	166193	3.968	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	8.174	79	16212	3.321	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.400%
46) 2-Hexanone-d5	8.628	63	91106	47.230	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.460%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	45933	4.397	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	56662	4.531	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.600%
Target Compounds						
10) 1,1,2-Trichloro-1,2,2-...	2.576	101	38057	3.416	ug/L	99
13) Acetone	2.641	43	13680	9.191	ug/L	96
34) Trichloroethene	6.538	95	9444	0.712	ug/L	93
47) Tetrachloroethene	8.547	164	28831	2.891	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
Data File : VU061630.D
Acq On : 13 Nov 2024 17:41
Operator : MD/SY
Sample : P4747-08
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCPC7

Quant Time: Nov 14 01:31:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 14 00:27:20 2024
Response via : Initial Calibration

