

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
 Data File : VU061654.D
 Acq On : 14 Nov 2024 04:16
 Operator : MD/SY
 Sample : P4803-18 5X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A0AS3

Quant Time: Nov 14 05:13:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 14 04:06:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	188947	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	171451	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	76538	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	50876	4.167	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.400%
7) Chloroethane-d5	1.911	69	44064	4.228	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.600%
11) 1,1-Dichloroethene-d2	2.554	65	24666	4.503	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.000%
20) 2-Butanone-d5	4.650	46	125744	48.527	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.060%
24) Chloroform-d	5.052	84	109014	4.492	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	5.692	65	57645	4.950	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%
32) Benzene-d6	5.714	84	232035	4.892	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
36) 1,2-Dichloropropane-d6	6.679	67	70976	5.074	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.891	98	198160	4.482	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	8.174	79	21851	4.240	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.643	63	111939	54.974	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.940%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	54108	4.906	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	67762	4.971	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds						
22) cis-1,2-Dichloroethene	4.656	96	3786	0.268	ug/L	96
33) Benzene	5.766	78	29956	0.577	ug/L	100
51) Chlorobenzene	9.441	112	72314	2.011	ug/L	97
65) 1,4-Dichlorobenzene	11.833	146	7000	0.283	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
Data File : VU061654.D
Acq On : 14 Nov 2024 04:16
Operator : MD/SY
Sample : P4803-18 5X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A0AS3

Quant Time: Nov 14 05:13:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 14 04:06:27 2024
Response via : Initial Calibration

