

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
 Data File : VU061620.D
 Acq On : 13 Nov 2024 12:43
 Operator : MD/SY
 Sample : VU1113WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK082

Quant Time: Nov 14 01:29:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 14 00:27:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	189052	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	177059	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	76967	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	63898	5.231	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.600%
7) Chloroethane-d5	1.911	69	56102	5.380	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.600%
11) 1,1-Dichloroethene-d2	2.563	65	28272	5.159	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.200%
20) 2-Butanone-d5	4.628	46	123910	47.793	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.580%
24) Chloroform-d	5.055	84	121876	5.020	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	5.692	65	62025	5.323	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.400%
32) Benzene-d6	5.717	84	260885	5.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.600%
36) 1,2-Dichloropropane-d6	6.682	67	77444	5.361	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.200%
41) Toluene-d8	7.891	98	234470	5.136	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.800%
43) trans-1,3-Dichloroprop...	8.174	79	25598	4.810	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.200%
46) 2-Hexanone-d5	8.627	63	107797	51.263	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.520%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	58526	5.139	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	75539	5.510	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
Data File : VU061620.D
Acq On : 13 Nov 2024 12:43
Operator : MD/SY
Sample : VU1113WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK082

Quant Time: Nov 14 01:29:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 14 00:27:20 2024
Response via : Initial Calibration

