

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
 Data File : VU061642.D
 Acq On : 13 Nov 2024 22:59
 Operator : MD/SY
 Sample : VU1113WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK083

Quant Time: Nov 14 04:07:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 14 04:06:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	178199	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	161126	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	69952	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	53548	4.651	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.000%
7) Chloroethane-d5	1.911	69	48707	4.955	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.200%
11) 1,1-Dichloroethene-d2	2.560	65	22656	4.386	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	4.628	46	128449	52.561	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.120%
24) Chloroform-d	5.052	84	108833	4.756	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
26) 1,2-Dichloroethane-d4	5.692	65	57914	5.273	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	5.718	84	224891	5.045	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
36) 1,2-Dichloropropane-d6	6.679	67	69887	5.316	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.400%
41) Toluene-d8	7.891	98	188492	4.537	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	8.174	79	23332	4.818	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.400%
46) 2-Hexanone-d5	8.627	63	106099	55.445	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.880%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	53325	5.145	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	62232	4.995	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111324\
Data File : VU061642.D
Acq On : 13 Nov 2024 22:59
Operator : MD/SY
Sample : VU1113WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK083

Quant Time: Nov 14 04:07:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 14 04:06:27 2024
Response via : Initial Calibration

