

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111424\
 Data File : VU061662.D
 Acq On : 14 Nov 2024 11:07
 Operator : MD/SY
 Sample : VIBLK162
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK162

Quant Time: Nov 15 00:43:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 15 00:40:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	189956	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	177348	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	77452	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	52868	4.308	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.911	69	46816	4.468	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.400%
11) 1,1-Dichloroethene-d2	2.563	65	23488	4.265	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.400%
20) 2-Butanone-d5	4.631	46	118066	45.322	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.640%
24) Chloroform-d	5.052	84	107928	4.424	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.695	65	54994	4.697	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	5.718	84	227548	4.638	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	6.682	67	68736	4.750	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.891	98	202176	4.421	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
43) trans-1,3-Dichloroprop...	8.174	79	22334	4.190	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.631	63	97139	46.119	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.240%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	50493	4.426	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	67191	4.871	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111424\
Data File : VU061662.D
Acq On : 14 Nov 2024 11:07
Operator : MD/SY
Sample : VIBLK162
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK162

Quant Time: Nov 15 00:43:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 15 00:40:52 2024
Response via : Initial Calibration

